

Performance of the Avioq VioOne™ HIV Profile™ Supplemental Assay for Confirmation and Differentiation of HIV-1 and HIV-2 Antibodies

***2022 Advancing HIV, STI and Viral Hepatitis
Testing Conference***



March 30, 2022

Disclaimer

Berry Bennett, Sally Fordan, Jessica Ribbing, Susanne Crowe, Marie-Claire Rowlinson

*Florida Department of Health, Bureau of Public Health Laboratories,
Jacksonville, FL., USA*

Disclaimer: Tradenames are used for informational purposes and do not constitute an endorsement by the Florida Department of Health

Disclosure: I do not have any financial relationship with commercial entities to disclose.




Background

- The Florida Bureau of Public Health Laboratories (BPHL) serves 67 county health departments (CHD) and contracted community-based organizations (CBO) statewide.
- The BPHL performs >90,000 HIV-1/2 diagnostic tests per year with approximately 2,800 specimens reflex to HIV-1/2 supplemental Ab testing. Of these approximately 89% are confirmed HIV-1 Ab positive, 9% HIV-1/2 Ab negative and the remaining 2% Ab indeterminate.



2014 HIV Diagnostic Algorithm¹




Avioq VioOne™ HIV Profile™ Supplemental Assay

- ELISA microplate platform
 - Four HIV-1 antigens (p24, p65, gp41 and gp160)
 - One HIV-2 antigen (gp36)
- 96-well microplate washer/reader required
- An HIV-1/2 Positive control and a Negative control must be included in each run with up to 10 clinical specimens per plate
- FDA approved Oct 2020, intended for serum, plasma



Avioq VioOne™ HIV Profile™ Supplemental Assay



- Specimen wells with $S/CO \geq 1.0$ are reactive
- HIV-1 positive if $S/CO \geq 1.0$ for 2 or more HIV-1 wells
(HIV-1 indeterminate if only 1 antigen well)
- HIV-2 positive if $S/CO \geq 1.0$ for the HIV-2 gp36 well


Bio-Rad Geenius™ HIV-1/2 Supplemental Assay

- Single-use immunochromatographic assay for the confirmation and differentiation of antibodies to HIV-1 and HIV-2
- Detects IgG antibodies
 - Two HIV-2 bands (gp36, gp140)
 - Four HIV-1 bands (p31, gp160, p24, gp41)
 - Internal control band w/each cassette



1 = gp36
2 = gp140
3 = p31
4 = gp160
5 = p24
6 = gp41
C = Control

- Proprietary Geenius software and reader
- FDA approved October 2014, with mandatory software upgrade (APF V2.0) Oct 2020
- Intended use: whole blood, serum, plasma



HIV-1/2 Ab Supplemental Interpretations

| Geenius Final Assay Interpretation | VioOne HIV Profile Result Interpretation |
|--|---|
| HIV (Antibody) Negative | HIV Negative |
| HIV-1 Indeterminate | HIV-1 Indeterminate |
| HIV-2 Indeterminate | No Equivalent |
| HIV Indeterminate | No Equivalent |
| HIV-1 Positive | HIV-1 Positive |
| HIV-2 Positive | HIV-2 Positive |
| HIV-2 Positive with HIV-1 Cross Reactivity | HIV-2 Positive with Reactivity to HIV-1 Antigens HIV-1 gp41 S/CO ≤ HIV-2 gp36 S/CO |
| No Equivalent | HIV-1 Positive with Reactivity to HIV-2 Antigen HIV-1 gp41 S/CO > HIV-2 gp36 S/CO |
| HIV Positive Untypable | No Equivalent |

Study Design

- Limited study (n=59) on the performance of the Profile Supplemental Assay
 - 44 in-house remnant serum/plasma
 - 15-member (7 HIV-1 reactive, 7 HIV-2 reactive and 1 HIV-1/2 nonreactive) antibody performance panel*
- In-house clinical specimen eligibility criteria:
 - Each characterized as one of 6 different Geenius “assay interpretations”
 - No more than one freeze-thaw cycle
 - Specimens deidentified and tested in singlet, minimum of 200ul for Profile and 20ul for Geenius
 - Clinical specimens routinely received for serostatus determination, April – August 2021
- Profile and Geenius assays performed in batches of 10 or less clinical specimens within 24 hours between August 3 – 18, 2021

*SeraCare HIV-1/2 Mixed Titer Accuset Performance Panel
(Jan 2015)

Results: HIV Seropositive Specimens, n=43

| “Assay Interpretations” | | Profile HIV-1/HIV-2 | |
|-------------------------|----------------------|---------------------|--------------------|
| Geenius HIV-1/HIV-2 | HIV-1 Positive | HIV-1 Positive (36) | HIV-2 Positive (7) |
| | HIV-1 Indeterminate | 2 ^a | |
| | HIV Untypable | 1 | |
| | HIV Ab Negative | | |
| | HIV-2 Positive | | 7 |
| | HIV-2 Indeterminate | | |
| | Positive % Agreement | | 95.3% (41/43) |

Results based on Geenius IFU (APF V2.0) and Profile IFU “assay interpretations”.

- Profile identified 100% (43/43) as HIV seropositive w/o reflex HIV RNA testing.
- Geenius identified 95.3% (41/43) as HIV seropositive w/o reflex HIV RNA testing.

^a Evidence of past HIV-1 diagnosis, in care

Results: Discordant HIV Specimens, n=16

| “Assay Interpretations” | | Profile HIV-1/HIV-2 | <p>Results based on Geenius IFU (APF V2.0) and Profile IFU “assay interpretations” .</p> <ul style="list-style-type: none">• All discordant algorithm specimens reflexed to HIV RNA testing. |
|-------------------------|-----------------|---------------------|--|
| HIV-1 Positive | | | |
| HIV-1 Indeterminate | 3 ^a | | |
| HIV Untypable | | | |
| HIV Ab Negative | 12 ^b | | |
| HIV-2 Positive | | | |
| HIV-2 Indeterminate | 1 ^c | | |
| % Agreement | | 100% (16/16) | |

Conclusion

- Number of specimens requiring RNA testing to resolve serostatus,
Profile = 27% (16/59)
Geenius = 30.5% (18/59)
- Both assays provided same day reporting, up to 10 laboratory results available within 180 minutes (w/approx. 33% hands-on) with Profile and 45 minutes (100% hands-on) with Geenius.
- Both assays performed well in our PH laboratory setting, however technical staff with microplate immunoassay experience is recommended.



Discussion

- This study identified (13) Geenius HIV-1 Ab reactive/HIV-2 Ab indeterminate “analyte interpretations”. Laboratories are cautioned NOT to report “analyte interpretations”. *“The final assay interpretation should always be reported to the ordering healthcare provider” – Bio-Rad Geenius HIV-1/2 APF (Assay Protocol File) V1.3.*
- More data is needed to assess continued use of HIV-1/HIV-2 antibody differentiating assays in contrast with HIV-1/HIV-2 qualitative RNA.
- HIV-1/HIV-2 antibody assay results contribute to an algorithm-defined acute HIV-1 infection, regardless of its position in the algorithm design.
- Manual microplate washing and multi-channel pipetting is labor intensive, Avioq, Inc is evaluating robotic pipetting and/or automated washing options.

Limitations

- Only the Geenius repeatedly HIV-2 Indeterminate “assay interpretation” case reflexed to HIV-2 RNA.
- HIV-2 RNA testing was NOT performed on the thirteen (13) Geenius HIV-1 positive, HIV-2 indeterminate “analyte interpretations”.
- This study was not intended to assess the sensitivity and specificity of the HIV-1/2 Profile Supplemental Assay.

Acknowledgements

BPHL

Jessica Ribbing
Sally Fordan
Susanne Crowe
Mary Ritchie
Marie-Claire Rowlinson

Avioq Inc.

Don Lockwood
Mike Cronin
Brandi Le
Sandy Goldman
Tim Holody

Thank you & Questions

Berry.Bennett@FLHealth.gov

904-791-1527

